Test samples RAYCON

Used for RAYCON

■ XRay Test Card Single V2A

Article number (Ball sizes)		Material	Example
77187830	(0.3 mm)	Stainless steel 1.4301 Density: 8.0 g/cm³	
77187831	(0.4 mm)		
77187833	(0.5 mm)		
77187834	(0.6 mm)		
77187835	(0.8 mm)		
77187836	(1.0 mm)		
77187837	(1.2 mm)		
77187838	(1.5 mm)		
77187839	(1.8 mm)		
77187840	(2.0 mm)		
77187841	(2.5)mm		
77187842	(3.0 mm)		

■ XRay Test Card Single Glas

77187853	(0.8 mm)	Glass Density: 2.5 g/cm³	
77187854	(1.0 mm)		
77187855	(1.2 mm)		
77187856	(1.5 mm)		
77187857	(2.0 mm)		
77187858	(2.5 mm)		
77187859	(3.0 mm)		
77187860	(3.5 mm)		
77187861	(4.0 mm)		
77187862	(5.0 mm)		
77187863	(6.0 mm)		
77187864	(7.0 mm)		
77187865	(8.0 mm)		
77187867	(9.0 mm)		
77187868	(10.0 mm)		

■ XRay Test Card Multi V2A

Article number	Material	Ball Ø [mm]	Example
77187843	Stainless Steel 1.4301 Density: 8.0 g/cm³	Ø 0.3 mm Ø 0.4 mm Ø 0.5 mm Ø 0.6 mm Ø 0.8 mm	
77187844	Stainless Steel 1.4301 Density: 8.0 g/cm³	Ø 0.8 mm Ø 1.0 mm Ø 1.2 mm Ø 1.5 mm Ø 1.8 mm	
77187845	Stainless Steel 1.4301 Density: 8.0 g/cm³	Ø 1.8 mm Ø 2.0 mm Ø 2.5 mm Ø 3.0 mm Ø 3.5 mm	

Mobergs Produktkontroll AB • Fabriksvägen 2 • SE-245 34 Staffanstorp
Tel. +46 (0)40 6808060 • info@mobergs.se • www.mobergs.se

Test samples RAYCON

XRay Test Card Multi Glas

77187846	Glass Density: 2.5 g/cm³	Ø 0.8 mm Ø 1.0 mm Ø 1.2 mm Ø 1.5 mm Ø 2.0 mm	
77187847	Glass Density: 2.5 g/cm³	Ø 2.0 mm Ø 2.5 mm Ø 3.0 mm Ø 3.5 mm Ø 4.0 mm	
77187848	Glass Density: 2.5 g/cm³	Ø 4.0 mm Ø 5.0 mm Ø 6.0 mm Ø 7.0 mm Ø 8.0 mm	

XRay Test Card Multi Wire V4A

Article number	Material	Ball Ø [mm]	Example
77187869	Stainless steel wire 1.4401 Density: 8.0g/cm³	Ø 0.3 x 2.0 mm Ø 0.4 x 2.0 mm Ø 0.5 x 2.0 mm Ø 0.6 x 2.0 mm Ø 0.8 x 2.0 mm	
77187870	Stainless steel wire 1.4401 Density: 8.0g/cm³	Ø 0.3 x 5.0 mm Ø 0.4 x 5.0 mm Ø 0.5 x 5.0 mm Ø 0.6 x 5.0 mm Ø 0.8 x 5.0 mm	

XRay Test Card Multi Ceramic

77187849	Ceramics Density: 3.9g/cm³	Ø 2.0 mm Ø 3.0 mm Ø 4.0 mm Ø 5.0 mm Ø 6.0 mm	

XRay Test Card Multi EPDM

77187850	EPDM Density: 1.2g/cm³	Ø 6.0 mm Ø 7.0 mm Ø 8.0 mm Ø 9.0 mm Ø 10.0 mm	

XRay Test Card Multi PTFE

77187851	PTFE Density: 2.1g/cm³	Ø 3.0 mm Ø 4.0 mm Ø 5.0 mm Ø 6.0 mm Ø 7.0 mm	

Mobergs Produktkontroll AB • Fabriksvägen 2 • SE-245 34 Staffanstorp
Tel. +46 (0)40 6808060 • info@mobergs.se • www.mobergs.se

Test samples RAYCON

XRay Test Card Multi ALU

77187851	Aluminium Density: 2.7g/cm³	Ø 2.0 mm Ø 3.0 mm Ø 4.0 mm Ø 5.0 mm Ø 6.0 mm	

Test XRay Dome Bulk V2A

Article number	(Ball sizes)	Material	Example
77187880	(0.4 mm)	Stainless Steel 1.4301 Density: 8.0g/cm³ Diameter Ø 25 mm	
77187881	(0.5 mm)		
77187882	(0.6 mm)		
77187883	(0.8 mm)		
77187884	(1.0 mm)		
77187886	(1.2 mm)		
77187887	(1.5 mm)		
77187888	(1.8 mm)		
77187889	(2.0 mm)		
77187890	(2.5)mm		

Test XRay Dome Bulk Glas

Article number		Material	Example
77187891	(1.5 mm)	Glass Density: 2.5 g/cm³ Diameter Ø 25 mm	
77187892	(2.0 mm)		
77187893	(2.5 mm)		
77187894	(3.0 mm)		
77187895	(3.5 mm)		
77187896	(4.0 mm)		

Test XRay Stick Bone

Article number		Material	Example
77187871	(2x2x2 mm)	synthetic Bone Density: 1.6g/cm³ Possibly less	
77187872	(3x3x3 mm)		
77187873	(4x4x4 mm)		
77187874	(6x6x6 mm)		
77187875	(8x8x8 mm)		
77187876	(10x10x10 mm)		


XRay Test Card Multi Bone

Article number	Material	Ball Ø [mm]	Example
77187878	synthetic Bone Density: 1.6g/cm³	2x2x2 mm 3x3x3 mm 4x4x4 mm 5x5x5 mm 6x6x6 mm	
77187877	synthetic Bone Density: 1.6g/cm³	6x6x6 mm 7x7x7 mm 8x8x8 mm 9x9x9 mm 10x10x10 mm	


Mobergs Produktkontroll AB • Fabriksvägen 2 • SE-245 34 Staffanstorp
Tel. +46 (0)40 6808060 • info@mobergs.se • www.mobergs.se

Test samples RAYCON

XRay Test Card Area V2A

Article number		Material	Example
33012306 33012308 33012344 33012382 33012412 33012452	(0.6 mm) (0.8 mm) (1.0 mm) (1.2 mm) (1.5 mm) (1.8 mm)	Stainless steel 1.4301 Density: 8.0g/cm³	

XRay Test Card Area Glas

Article number		Material	Example
33012490 33012544 33012546 33012602 33012776 33012838	(1,5 mm) (2,0 mm) (2,5 mm) (3,0 mm) (3,5 mm) (4,0 mm)	Glass Density: 2.5 g/cm³	

Declaration of conformity (ISO/IEC 17050-1)

- Factory certificate 2.1 of the test sample according to EN 10204
- Foodstuffs certificate of compliance for test sample sheath
  - FDA declaration of conformity for the sheathing of the test sample
  - (EC) No. 1935/2004*
    - Declaration of Food Conformity according to REGULATION (EU) No. 10/2011 (on request)

The Sesotec certificates of the test cards and test samples are available for download in digital form on the Mobergs website and are included in the price.

For area test cards, the FDA certificate is available on request.

*** except test cards**

Mobergs Produktkontroll AB • Fabriksvägen 2 • SE-245 34 Staffanstorp
Tel. +46 (0)40 6808060 • info@mobergs.se • www.mobergs.se