Sesotec GmbH
Regener Straße 130
D-94513 Schönberg
Deutschland
Telefon: 08554 3080
Fax: 08554 2606
info@sesotec.com
www.sesotec.com

Konformitätserklärung (ISO/IEC 17050-1)
Testkörper

Supplier's declaration of conformity (in accordance with ISO/IEC 17050-1)
Test sample

Werkstoff Material	**Durchmesser** Diameter	**Ident-Nr.** Ident-no.	**Dichte** Density	**Maß-/Formgenauigkeit** Dimensional accuracy DIN 5401-2002-08	**Magnetische Permeabilität** Magnetic permeability
Ceramic	2,00 mm	IK80093	3,88 g/cm³	Grade 10	--
	2,00 mm	IK80101		Grade 10	--
	3,00 mm	IK80096		Grade 10	--
	4,00 mm	IK80094		Grade 10	--
		IK80099		Grade 10	--
	5,00 mm	IK80091		Grade 10	--
		IK80100		Grade 10	--
	6,00 mm	IK80095	3,90 g/cm³	Grade 10	--

Lebensmittelbescheinigung Testkörperummantelung
Food certificate test ball cover

☒ FDA-Konformitätserklärung / FDA-Declaration of conformity

Hiermit bestätigen wir, dass die Ummantelung unserer Testkörper aus Kunststoffen/Elastomeren, die bei bestimmungsgemäßem Gebrauch in Kontakt mit Lebensmitteln kommen können, die allgemeinen Anforderungen der FDA-Richtlinie 21 CFR §177 erfüllen.

We hereby certify that our test sample covers made of plastics/elastomers, which could come into contact with foodstuffs in the course of their intended use, meet the requirements under FDA directive 21 CFR Section 177

Enthaltene Additive entsprechen jeweils den zutreffenden FDA-Anforderungen.
Any additives that they contain meet the respectively applicable FDA requirements:

POM Blau / POM Blue (TestSticks, Testballs) - CFR 21, Part 177.2470 (Polyoxymethylene copolymer) - CFR 21, Part 178.3297 (colorants)	PE / PE (FlexStick Schlauch/FlexStick hose) - CFR 21, Part 177.1520 (Olefin polymers)	EVA Co-polymer (Flyball) - CFR 21, Part 177.1350 (EVA Co-polymers)
CARD LINE XD (Xray-Cards/Teststripes) - CFR 21, Part 177.1580 (Polycarbonate)		

Sesotec GmbH
10.01.2024

i. V. Dieter Sprenzinger
- Manager Quality -